

www.jantek.com.tw

## **JTT-700**

## Temperature Measuring Device



### 

MIFARE READER / MIFARE MODULE / FINGERPRINT READER

Taiwan Jantek Electronics Ltd. TIME AND ATTENDANCE SYSTEM / ELECTRIC DOOR BOLT ACCESSORIES / COPIER CONTROLLER / PUSH BUTTON

Tel: 886-2-2299-1255(Rep.) | Fax: 886-2-2299-1257 / 2299-7412 6F-1, No. 1, Wuquan 1St.Rd., Xinzhuang District, New Taipei City 24892, Taiwan,

# JTT-700 Temperature Measuring Device

#### Features

- · This device and be used at entrance to detect people's body temperature.
- · Abnormal body temperature is settable. Value is from 0.0°C to 99.9°C. Default value is 37.3°C. When detecting body temperature higher than set value, display will show temperature in red and voice warning will occur.
- · A built-in relay output could be used to control access.
- $\cdot$  A RS-485 output could be used to upload temperature data to software for application.
- · Reads in Celsius or Fahrenheit.
- · Temperatures vary by approximately ±0.2°C
- · Prop stand is optional.

### Specifications

- · 4 digits display.
- · Adjustable normal body temperature range.
- · 12 selectable languages. Including Mandarin, English, French, Korean, Japanese, German, Italian, Spanish, Russian, Portuguese, Arabic and Hindi
- · Has °C/°F key and volume key.
- · Has lock key, system status key and mute key.
- · Has 1 RS232 contact, 1 RS485 contact and 1 internet contact (optional)
- · Has 1 relay output
- · Dimension: 240\*150\*28mm. Prop stand height is adjustable from 70 to 210 cm.
- Power supply DC12V/2A

### Temperature parameter

- · Measuring speed is less than 0.1 second.
- · Temperatures vary by approximately ±0.2°C
- · Maximum reading range is 300 cm.
- · Best reading range is 20 to 200 cm.
- · Best environment temperature while measuring is 10 to 35°C
- · Reads in Celsius or Fahrenheit.